

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031723\
 Data File : VX034606.D
 Acq On : 18 Mar 2023 02:55
 Operator : JC/MD
 Sample : 01932-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 JPP-6A-031423

Quant Time: Mar 20 00:14:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	231844	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.763	114	396986	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	367811	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	161894	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	162396	48.148	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.300%
35) Dibromofluoromethane	5.385	113	124508	47.502	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.000%
50) Toluene-d8	8.647	98	490756	50.155	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.320%
62) 4-Bromofluorobenzene	11.079	95	189696	47.997	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.000%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	77572	48.825	ug/l	96
20) Methylene Chloride	2.788	84	18527	6.008	ug/l	96
37) Ethyl Acetate	4.721	43	59164	13.452	ug/l	98
43) Isopropyl Acetate	6.342	43	243950	32.742	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031723\
Data File : VX034606.D
Acq On : 18 Mar 2023 02:55
Operator : JC/MD
Sample : 01932-14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
JPP-6A-031423

Quant Time: Mar 20 00:14:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

