

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031723\
 Data File : VX034613.D
 Acq On : 18 Mar 2023 05:40
 Operator : JC/MD
 Sample : 01963-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 NYE-SOIL-0316

Quant Time: Mar 20 00:17:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	222792	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	380943	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	370133	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	177015	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	155376	47.938	ug/l	0.01
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.880%
35) Dibromofluoromethane	5.391	113	121121	48.156	ug/l	0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.320%
50) Toluene-d8	8.647	98	470057	50.063	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.120%
62) 4-Bromofluorobenzene	11.079	95	224501	59.195	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	118.400%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	36043	23.608	ug/l	97
20) Methylene Chloride	2.788	84	18016	6.080	ug/l	95
37) Ethyl Acetate	4.721	43	29091	6.893	ug/l	98
43) Isopropyl Acetate	6.342	43	125916	17.612	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031723\
Data File : VX034613.D
Acq On : 18 Mar 2023 05:40
Operator : JC/MD
Sample : 01963-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
NYE-SOIL-0316

Quant Time: Mar 20 00:17:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

