

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031723\
 Data File : VX034614.D
 Acq On : 18 Mar 2023 06:04
 Operator : JC/MD
 Sample : 01964-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 UNION-2

Quant Time: Mar 20 00:17:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	224916	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	380998	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	358010	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	181444	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	156363	47.787	ug/l	0.01
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.580%
35) Dibromofluoromethane	5.385	113	118817	47.233	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.460%
50) Toluene-d8	8.647	98	472120	50.276	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.560%
62) 4-Bromofluorobenzene	11.079	95	217311	57.291	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	114.580%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	48121	31.221	ug/l	99
20) Methylene Chloride	2.788	84	11797	3.944	ug/l	94
37) Ethyl Acetate	4.721	43	31749	7.522	ug/l	98
43) Isopropyl Acetate	6.336	43	143570	20.078	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031723\
Data File : VX034614.D
Acq On : 18 Mar 2023 06:04
Operator : JC/MD
Sample : 01964-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
UNION-2

Quant Time: Mar 20 00:17:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

