



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031723\
Data File : VX034573.D
Acq On : 17 Mar 2023 09:32
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Title : SW846 8260
Last Update : Mon Mar 13 10:29:32 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	21.8	18845	PASS
75	95	30	60	53.5	46376	PASS
95	95	100	100	100.0	86629	PASS
96	95	5	9	7.0	6075	PASS
173	174	0.00	2	1.1	670	PASS
174	95	50	100	73.3	63541	PASS
175	174	5	9	7.2	4560	PASS
176	174	95	101	97.6	62024	PASS
177	176	5	9	6.6	4091	PASS