

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031725\
 Data File : VX045299.D
 Acq On : 17 Mar 2025 12:02
 Operator : JC/MD
 Sample : VX0317WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0317WBL01

Quant Time: Mar 18 01:23:06 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	73730	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	144901	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	130224	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	50695	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63453	54.105	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	108.200%
35) Dibromofluoromethane	5.379	113	50888	52.521	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.040%
50) Toluene-d8	8.647	98	178851	50.916	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.840%
62) 4-Bromofluorobenzene	11.079	95	60611	52.072	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	104.140%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031725\
Data File : VX045299.D
Acq On : 17 Mar 2025 12:02
Operator : JC/MD
Sample : VX0317WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0317WBL01

Quant Time: Mar 18 01:23:06 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

