

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031725\
 Data File : VX045306.D
 Acq On : 17 Mar 2025 15:25
 Operator : JC/MD
 Sample : PB167134ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB167134ZHE#03

Quant Time: Mar 18 01:28:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	75824	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	148947	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	138042	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	55646	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	66442	55.089	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 110.180%	
35) Dibromofluoromethane	5.385	113	52563	52.776	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.560%	
50) Toluene-d8	8.647	98	188568	52.224	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 104.440%	
62) 4-Bromofluorobenzene	11.079	95	64396	53.821	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 107.640%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	34573	61.783	ug/l	99
18) Methyl Acetate	2.703	43	2848	2.116	ug/l	94
20) Methylene Chloride	2.782	84	3300	2.977	ug/l	96
25) 2-Butanone	4.574	43	16544	19.641	ug/l	99
43) Isopropyl Acetate	6.342	43	78229	30.059	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031725\
Data File : VX045306.D
Acq On : 17 Mar 2025 15:25
Operator : JC/MD
Sample : PB167134ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB167134ZHE#03

Quant Time: Mar 18 01:28:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

