

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031725\
 Data File : VX045309.D
 Acq On : 17 Mar 2025 16:35
 Operator : JC/MD
 Sample : Q1568-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 JC-03-03132025

Quant Time: Mar 18 01:30:25 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	72781	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	144174	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	136750	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	53431	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64867	56.031	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 112.060%	
35) Dibromofluoromethane	5.379	113	49986	51.850	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.700%	
50) Toluene-d8	8.647	98	185557	53.092	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 106.180%	
62) 4-Bromofluorobenzene	11.079	95	63019	54.414	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 108.820%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.977	59	4623	21.082	ug/l #	92
16) Acetone	2.380	43	26005	48.415	ug/l	100
18) Methyl Acetate	2.703	43	2695	2.086	ug/l	92
20) Methylene Chloride	2.782	84	3081	2.896	ug/l	94
25) 2-Butanone	4.568	43	18901	23.377	ug/l	95
43) Isopropyl Acetate	6.342	43	87360	34.678	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031725\
Data File : VX045309.D
Acq On : 17 Mar 2025 16:35
Operator : JC/MD
Sample : Q1568-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
JC-03-03132025

Quant Time: Mar 18 01:30:25 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

