

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031725\
 Data File : VX045310.D
 Acq On : 17 Mar 2025 16:58
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 18 01:30:57 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	101	0.00
2 T	Dichlorodifluoromethane	0.629	0.591	6.0	96	0.00
3 P	Chloromethane	0.770	0.677	12.1	91	0.00
4 C	Vinyl Chloride	0.764	0.622	18.6#	83	0.00
5 T	Bromomethane	0.300	0.267	11.0	93	0.00
6 T	Chloroethane	0.352	0.339	3.7	92	0.00
7 T	Trichlorofluoromethane	1.011	0.942	6.8	91	0.00
8 T	Diethyl Ether	0.396	0.346	12.6	91	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.581	0.618	-6.4	106	0.00
10 T	Methyl Iodide	0.726	0.729	-0.4	96	0.00
11 T	Tert butyl alcohol	0.151	0.139	7.9	105	0.00
12 CM	1,1-Dichloroethene	0.614	0.587	4.4#	96	0.00
13 T	Acrolein	0.174	0.054	69.0#	32#	0.00
14 T	Allyl chloride	1.094	1.138	-4.0	104	0.00
15 T	Acrylonitrile	0.404	0.409	-1.2	101	0.00
16 T	Acetone	0.369	0.365	1.1	105	0.00
17 T	Carbon Disulfide	1.636	1.265	22.7	77	0.00
18 T	Methyl Acetate	0.888	1.107	-24.7	130	0.00
19 T	Methyl tert-butyl Ether	2.061	2.117	-2.7	103	0.00
20 T	Methylene Chloride	0.731	0.692	5.3	101	0.00
21 T	trans-1,2-Dichloroethene	0.607	0.583	4.0	94	0.00
22 T	Diisopropyl ether	2.224	2.267	-1.9	100	0.00
23 T	Vinyl Acetate	1.856	1.947	-4.9	100	0.00
24 P	1,1-Dichloroethane	1.247	1.233	1.1	101	0.00
25 T	2-Butanone	0.555	0.578	-4.1	101	0.00
26 T	2,2-Dichloropropane	0.585	0.879	-50.3#	149	0.00
27 T	cis-1,2-Dichloroethene	0.749	0.729	2.7	97	0.00
28 T	Bromochloromethane	0.594	0.604	-1.7	102	0.00
29 T	Tetrahydrofuran	0.373	0.368	1.3	100	0.00
30 C	Chloroform	1.239	1.242	-0.2#	101	0.00
31 T	Cyclohexane	1.082	1.040	3.9	95	0.00
32 T	1,1,1-Trichloroethane	1.000	1.016	-1.6	101	0.00
33 S	1,2-Dichloroethane-d4	0.795	0.828	-4.2	111	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	99	0.00
35 S	Dibromofluoromethane	0.334	0.350	-4.8	106	0.00
36 T	1,1-Dichloropropene	0.459	0.457	0.4	97	0.00
37 T	Ethyl Acetate	0.591	0.585	1.0	96	0.00
38 T	Carbon Tetrachloride	0.465	0.476	-2.4	101	0.00
39 T	Methylcyclohexane	0.549	0.583	-6.2	99	0.00
40 TM	Benzene	1.448	1.431	1.2	95	0.00
41 T	Methacrylonitrile	0.319	0.347	-8.8	101	0.00
42 TM	1,2-Dichloroethane	0.521	0.550	-5.6	103	0.00
43 T	Isopropyl Acetate	0.874	0.962	-10.1	103	0.00
44 TM	Trichloroethene	0.340	0.331	2.6	96	0.00
45 C	1,2-Dichloropropane	0.372	0.380	-2.2#	102	0.00
46 T	Dibromomethane	0.263	0.272	-3.4	100	0.00
47 T	Bromodichloromethane	0.516	0.544	-5.4	103	0.00
48 T	Methyl methacrylate	0.438	0.488	-11.4	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031725\
 Data File : VX045310.D
 Acq On : 17 Mar 2025 16:58
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 18 01:30:57 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.009	0.0	100	0.00
50 S	Toluene-d8	1.212	1.232	-1.7	101	0.00
51 T	4-Methyl-2-Pentanone	0.587	0.650	-10.7	106	0.00
52 CM	Toluene	0.849	0.868	-2.2#	96	0.00
53 T	t-1,3-Dichloropropene	0.431	0.498	-15.5	105	0.00
54 T	cis-1,3-Dichloropropene	0.503	0.565	-12.3	105	0.00
55 T	1,1,2-Trichloroethane	0.350	0.356	-1.7	99	0.00
56 T	Ethyl methacrylate	0.532	0.612	-15.0	107	0.00
57 T	1,3-Dichloropropane	0.605	0.623	-3.0	100	0.00
58 T	2-Chloroethyl Vinyl ether	0.259	0.309	-19.3	108	0.00
59 T	2-Hexanone	0.425	0.474	-11.5	105	0.00
60 T	Dibromochloromethane	0.366	0.379	-3.6	98	0.00
61 T	1,2-Dibromoethane	0.349	0.361	-3.4	101	0.00
62 S	4-Bromofluorobenzene	0.402	0.448	-11.4	111	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	101	0.00
64 T	Tetrachloroethene	0.320	0.303	5.3	94	0.00
65 PM	Chlorobenzene	1.058	1.070	-1.1	99	0.00
66 T	1,1,1,2-Tetrachloroethane	0.352	0.361	-2.6	102	0.00
67 C	Ethyl Benzene	1.843	1.910	-3.6#	98	0.00
68 T	m/p-Xylenes	0.675	0.697	-3.3	97	0.00
69 T	o-Xylene	0.681	0.694	-1.9	99	0.00
70 T	Styrene	1.101	1.175	-6.7	100	0.00
71 P	Bromoform	0.266	0.280	-5.3	102	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
73 T	Isopropylbenzene	3.903	4.028	-3.2	102	0.00
74 T	N-amyl acetate	1.831	2.054	-12.2	109	0.00
75 P	1,1,2,2-Tetrachloroethane	1.432	1.427	0.3	105	0.00
76 T	1,2,3-Trichloropropane	1.164	1.194	-2.6	109	0.00
77 T	Bromobenzene	0.921	0.899	2.4	100	0.00
78 T	n-propylbenzene	4.409	4.741	-7.5	103	0.00
79 T	2-Chlorotoluene	2.808	2.825	-0.6	103	0.00
80 T	1,3,5-Trimethylbenzene	3.156	3.282	-4.0	101	0.00
81 T	trans-1,4-Dichloro-2-butene	0.340	0.398	-17.1	123	0.00
82 T	4-Chlorotoluene	3.060	3.231	-5.6	104	0.00
83 T	tert-Butylbenzene	3.242	3.367	-3.9	104	0.00
84 T	1,2,4-Trimethylbenzene	3.175	3.313	-4.3	103	0.00
85 T	sec-Butylbenzene	3.884	4.220	-8.7	105	0.00
86 T	p-Isopropyltoluene	3.139	3.415	-8.8	104	0.00
87 T	1,3-Dichlorobenzene	1.643	1.682	-2.4	105	0.00
88 T	1,4-Dichlorobenzene	1.662	1.642	1.2	102	0.00
89 T	n-Butylbenzene	2.703	3.088	-14.2	107	0.00
90 T	Hexachloroethane	0.568	0.589	-3.7	102	0.00
91 T	1,2-Dichlorobenzene	1.648	1.658	-0.6	104	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.279	0.298	-6.8	108	0.00
93 T	1,2,4-Trichlorobenzene	0.938	0.960	-2.3	100	0.00
94 T	Hexachlorobutadiene	0.385	0.400	-3.9	107	0.00
95 T	Naphthalene	3.605	3.766	-4.5	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031725\
Data File : VX045310.D
Acq On : 17 Mar 2025 16:58
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Mar 18 01:30:57 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.990	1.004	-1.4	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6