


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031725\
Data File : VX045296.D
Acq On : 17 Mar 2025 10:38
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Title : SW846 8260
Last Update : Fri Feb 28 06:45:16 2025



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	22.6	8647	PASS
75	95	30	60	57.1	21874	PASS
95	95	100	100	100.0	38301	PASS
96	95	5	9	6.9	2638	PASS
173	174	0.00	2	0.6	164	PASS
174	95	50	100	73.7	28211	PASS
175	174	5	9	7.4	2082	PASS
176	174	95	101	98.9	27909	PASS
177	176	5	9	6.6	1843	PASS