

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031819\
 Data File : VX008174.D
 Acq On : 18 Mar 2019 18:59
 Operator : JC/SP
 Sample : K1909-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 97-98

Quant Time: Mar 19 12:05:03 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X031819W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Mon Mar 18 14:34:46 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.03	128	48200	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.86	114	263687	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.11	117	224609	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	6.07	65	123338	30.76	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	102.53%
60) 4-Bromofluorobenzene	11.13	95	102328	27.30	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	91.00%
63) Toluene-d8	8.71	98	325702	29.84	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	99.47%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX031819\
Data File : VX008174.D
Acq On : 18 Mar 2019 18:59
Operator : JC/SP
Sample : K1909-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
97-98

Quant Time: Mar 19 12:05:03 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X031819W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Mon Mar 18 14:34:46 2019
Response via : Initial Calibration

