

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031820\
 Data File : VX015328.D
 Acq On : 18 Mar 2020 15:57
 Operator : JC/SP
 Sample : VX0318WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0318WBL01

Quant Time: Mar 19 03:33:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 14:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	572229	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1074495	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	1001524	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	429675	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	349301	49.36	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.72%	
35) Dibromofluoromethane	5.48	113	333714	47.66	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.32%	
50) Toluene-d8	8.71	98	1264472	48.91	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.82%	
62) 4-Bromofluorobenzene	11.13	95	454317	46.20	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.40%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX031820\
Data File : VX015328.D
Acq On : 18 Mar 2020 15:57
Operator : JC/SP
Sample : VX0318WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0318WBL01

Quant Time: Mar 19 03:33:58 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 14:57:45 2020
Response via : Initial Calibration

