


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX031820\
 Data File : VX015320.D
 Acq On : 18 Mar 2020 11:49
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X031820W.M
 Title : SW846 8260
 Last Update : Wed Mar 18 14:57:45 2020



AutoFind: Scans 1679, 1680, 1681; Background Corrected with Scan 1673

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.9	10588	PASS
75	95	30	60	47.8	29903	PASS
95	95	100	100	100.0	62609	PASS
96	95	5	9	7.0	4369	PASS
173	174	0.00	2	0.5	190	PASS
174	95	50	100	65.1	40766	PASS
175	174	5	9	8.1	3306	PASS
176	174	95	101	95.3	38832	PASS
177	176	5	9	6.8	2658	PASS