

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031821\
 Data File : VX021276.D
 Acq On : 18 Mar 2021 11:51
 Operator : JC/MD
 Sample : VIBLK
 Misc : 25.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VIBLK

Quant Time: Mar 19 01:23:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXTR031821WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Mar 18 11:53:19 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.829	114	60508	5.00	ug/L	0.00
28) Chlorobenzene-d5	10.094	117	55124	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.059	152	26578	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.387	65	18820	4.97	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	99.40%
7) Chloroethane-d5	1.693	69	14029	4.68	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.60%
11) 1,1-Dichloroethene-d2	2.340	63	31898	3.61	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.20%
20) 2-Butanone-d5	4.523	46	51239	49.25	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.50%
24) Chloroform-d	5.141	84	38907	4.66	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.20%
26) 1,2-Dichloroethane-d4	6.035	65	22732	4.69	ug/L	0.01
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.80%
32) Benzene-d6	6.041	84	74494	5.06	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.20%
36) 1,2-Dichloropropane-d6	7.364	67	22905	4.96	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.20%
41) Toluene-d8	8.694	98	71370	4.87	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.40%
43) trans-1,3-Dichloroprop...	8.988	79	9940	4.80	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	96.00%
46) 2-Hexanone-d5	9.424	63	40833	50.00	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.00%
56) 1,1,2,2-Tetrachloroeth...	11.230	84	16626	4.70	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.00%
66) 1,2-Dichlorobenzene-d4	12.359	152	24209	4.90	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.00%
Target Compounds						
					Qvalue	
14) Carbon disulfide	2.546	76	1971	0.14	ug/L	100
16) Methylene chloride	2.829	84	1189	0.22	ug/L	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031821\
Data File : VX021276.D
Acq On : 18 Mar 2021 11:51
Operator : JC/MD
Sample : VIBLK
Misc : 25.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VIBLK

Quant Time: Mar 19 01:23:20 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXTR031821WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Mar 18 11:53:19 2021
Response via : Initial Calibration

