

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031822\
 Data File : VX027518.D
 Acq On : 18 Mar 2022 10:45
 Operator : JC/MD
 Sample : VX0318MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0318MBL01

Quant Time: Mar 19 03:04:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 15:02:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	333951	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	550793	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	489625	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	218893	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	207761	50.502	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.000%
35) Dibromofluoromethane	5.385	113	183085	51.665	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	103.320%
50) Toluene-d8	8.653	98	642904	47.943	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.880%
62) 4-Bromofluorobenzene	11.079	95	226038	45.860	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	91.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031822\
Data File : VX027518.D
Acq On : 18 Mar 2022 10:45
Operator : JC/MD
Sample : VX0318MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0318MBL01

Quant Time: Mar 19 03:04:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 15:02:59 2022
Response via : Initial Calibration

