

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031822\
 Data File : VX027529.D
 Acq On : 18 Mar 2022 15:43
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Mar 19 03:07:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 15:02:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	334706	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	559435	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	518917	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	270236	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	206134	49.993	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.980%
35) Dibromofluoromethane	5.391	113	185422	51.516	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	103.040%
50) Toluene-d8	8.653	98	663305	48.700	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.400%
62) 4-Bromofluorobenzene	11.085	95	253347	50.607	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	101.220%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	5374	3.198	ug/l	93
17) Carbon Disulfide	2.508	76	1916	0.242	ug/l	98
20) Methylene Chloride	2.794	84	2033	0.563	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031822\
Data File : VX027529.D
Acq On : 18 Mar 2022 15:43
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Mar 19 03:07:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 15:02:59 2022
Response via : Initial Calibration

