

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031822\
 Data File : VX027530.D
 Acq On : 18 Mar 2022 16:06
 Operator : JC/MD
 Sample : N1998-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-127

Quant Time: Mar 19 03:08:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 15:02:59 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	329619	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	550146	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	506828	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	249459	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	203128	50.024	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.040%	
35) Dibromofluoromethane	5.391	113	184839	52.221	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 104.440%	
50) Toluene-d8	8.653	98	646604	48.276	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 96.560%	
62) 4-Bromofluorobenzene	11.085	95	242463	49.251	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 98.500%	
Target Compounds						
						Qvalue
11) Tert butyl alcohol	2.965	59	34545	44.639	ug/l	98
16) Acetone	2.380	43	4816	2.910	ug/l	93
19) Methyl tert-butyl Ether	3.117	73	196351	19.014	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031822\
Data File : VX027530.D
Acq On : 18 Mar 2022 16:06
Operator : JC/MD
Sample : N1998-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-127

Quant Time: Mar 19 03:08:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 15:02:59 2022
Response via : Initial Calibration

