

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031822\
 Data File : VX027539.D
 Acq On : 18 Mar 2022 19:39
 Operator : JC/MD
 Sample : N1998-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-136

Quant Time: Mar 19 03:09:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 15:02:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	340212	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	569342	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	518748	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	233378	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	209261	49.930	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.860%
35) Dibromofluoromethane	5.391	113	187022	51.056	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.120%
50) Toluene-d8	8.653	98	668694	48.242	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.480%
62) 4-Bromofluorobenzene	11.085	95	236022	46.326	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	92.660%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.965	59	722897	905.050	ug/l	100
16) Acetone	2.380	43	5596	3.276	ug/l	98
19) Methyl tert-butyl Ether	3.117	73	5564073	522.026	ug/l	98
22) Diisopropyl ether	3.763	45	7259	0.752	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031822\
Data File : VX027539.D
Acq On : 18 Mar 2022 19:39
Operator : JC/MD
Sample : N1998-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-136

Quant Time: Mar 19 03:09:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 15:02:59 2022
Response via : Initial Calibration

