



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031822\
 Data File : VX027516.D
 Acq On : 18 Mar 2022 09:39
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Title : SW846 8260
 Last Update : Thu Mar 17 15:02:59 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.5	12423	PASS
75	95	30	60	49.5	35149	PASS
95	95	100	100	100.0	71077	PASS
96	95	5	9	6.8	4839	PASS
173	174	0.00	2	1.0	553	PASS
174	95	50	100	77.2	54843	PASS
175	174	5	9	7.7	4236	PASS
176	174	95	101	95.5	52363	PASS
177	176	5	9	6.6	3476	PASS