



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031822\
 Data File : VX027544.D
 Acq On : 18 Mar 2022 22:01
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Title : SW846 8260
 Last Update : Thu Mar 17 15:02:59 2022



AutoFind: Scans 1639, 1640, 1641; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.6	11483	PASS
75	95	30	60	48.2	35387	PASS
95	95	100	100	100.0	73416	PASS
96	95	5	9	6.7	4952	PASS
173	174	0.00	2	1.0	640	PASS
174	95	50	100	86.0	63163	PASS
175	174	5	9	7.1	4503	PASS
176	174	95	101	95.6	60395	PASS
177	176	5	9	6.4	3844	PASS