

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031824\
 Data File : VX040703.D
 Acq On : 18 Mar 2024 16:45
 Operator : JC/MD
 Sample : P1753-23
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VHBLK001

Quant Time: Mar 19 01:25:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML030724WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Mar 19 01:22:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	135232	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	127061	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	61694	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	63999	53.988	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	= 107.980%	
7) Chloroethane-d5	1.666	69	51891	54.830	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	= 109.660%	
11) 1,1-Dichloroethene-d2	2.306	65	28371	51.991	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 103.980%	
21) 2-Butanone-d5	4.452	46	90911	99.939	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	= 99.940%	
24) Chloroform-d	5.056	84	114202	51.876	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 103.760%	
26) 1,2-Dichloroethane-d4	5.958	65	84906	51.659	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 103.320%	
32) Benzene-d6	5.970	84	213124	54.672	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 109.340%	
36) 1,2-Dichloropropane-d6	7.306	67	64443	53.340	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	= 106.680%	
41) Toluene-d8	8.647	98	193629	51.542	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 103.080%	
43) trans-1,3-Dichloroprop...	8.952	79	31799	49.453	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 98.900%	
47) 2-Hexanone-d5	9.384	63	64989	97.489	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	= 97.490%	
56) 1,1,2,2-Tetrachloroeth...	11.189	84	92903	51.360	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	= 102.720%	
66) 1,2-Dichlorobenzene-d4	12.323	152	71599	55.411	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 110.820%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031824\
Data File : VX040703.D
Acq On : 18 Mar 2024 16:45
Operator : JC/MD
Sample : P1753-23
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VHBLK001

Quant Time: Mar 19 01:25:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML030724WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Mar 19 01:22:22 2024
Response via : Initial Calibration

