

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031824\
 Data File : VX040690.D
 Acq On : 18 Mar 2024 11:41
 Operator : JC/MD
 Sample : VX0318WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VBLK666

Quant Time: Mar 19 01:22:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML030724WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Mar 19 01:22:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.757	114	161350	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.049	117	153957	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	68141	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	67934	48.031	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	96.060%
7) Chloroethane-d5	1.666	69	53391	47.283	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	94.560%
11) 1,1-Dichloroethene-d2	2.307	65	30742	47.217	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.440%
21) 2-Butanone-d5	4.446	46	98002	90.295	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	90.300%
24) Chloroform-d	5.050	84	121623	46.304	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.600%
26) 1,2-Dichloroethane-d4	5.952	65	91498	46.659	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.320%
32) Benzene-d6	5.970	84	227374	48.138	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.280%
36) 1,2-Dichloropropane-d6	7.306	67	70277	48.007	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.020%
41) Toluene-d8	8.647	98	212633	46.713	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.420%
43) trans-1,3-Dichloroprop...	8.946	79	35958	46.152	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.300%
47) 2-Hexanone-d5	9.378	63	72925	90.283	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	90.280%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	101204	46.174	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.340%
66) 1,2-Dichlorobenzene-d4	12.317	152	70563	49.442	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.880%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031824\
Data File : VX040690.D
Acq On : 18 Mar 2024 11:41
Operator : JC/MD
Sample : VX0318WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VBLK666

Quant Time: Mar 19 01:22:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML030724WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Mar 19 01:22:22 2024
Response via : Initial Calibration

