

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031825\
 Data File : VX045323.D
 Acq On : 18 Mar 2025 14:27
 Operator : JC/MD
 Sample : Q1559-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT149S1-20250310

Quant Time: Mar 19 01:44:49 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	68678	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	132613	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	121917	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	44828	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	60251	55.153	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.300%
35) Dibromofluoromethane	5.379	113	46453	52.386	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.780%
50) Toluene-d8	8.647	98	166680	51.848	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.700%
62) 4-Bromofluorobenzene	11.079	95	54236	50.912	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	101.820%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	1327	2.618	ug/l #	85
19) Methyl tert-butyl Ether	3.111	73	32200	11.373	ug/l	97
44) Trichloroethene	7.129	130	4086	4.531	ug/l	88
64) Tetrachloroethene	9.275	164	891	1.141	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031825\
Data File : VX045323.D
Acq On : 18 Mar 2025 14:27
Operator : JC/MD
Sample : Q1559-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT149S1-20250310

Quant Time: Mar 19 01:44:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

