

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031825\
 Data File : VX045333.D
 Acq On : 18 Mar 2025 18:20
 Operator : JC/MD
 Sample : Q1559-19
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP05-20250312

Quant Time: Mar 19 01:48:55 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	67809	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	134986	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	124517	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	53305	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	61840	57.334	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 114.660%	
35) Dibromofluoromethane	5.385	113	47336	52.443	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.880%	
50) Toluene-d8	8.647	98	172527	52.723	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 105.440%	
62) 4-Bromofluorobenzene	11.079	95	59486	54.859	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 109.720%	
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.325	101	1222	1.551	ug/l	92
16) Acetone	2.380	43	1139	2.276	ug/l	98
44) Trichloroethene	7.123	130	11683	12.729	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031825\
Data File : VX045333.D
Acq On : 18 Mar 2025 18:20
Operator : JC/MD
Sample : Q1559-19
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP05-20250312

Quant Time: Mar 19 01:48:55 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

