

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031919\
 Data File : VX008185.D
 Acq On : 19 Mar 2019 12:23
 Operator : JC/SP
 Sample : IBLK
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Mar 20 07:22:59 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X031919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 11:34:11 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	315801	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	537325	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	456356	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	188216	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	234892	48.57	ug/l	0.00
Spiked Amount			Recovery	=	97.14%	
35) Dibromofluoromethane	5.50	113	169018	45.63	ug/l	0.00
Spiked Amount			Recovery	=	91.26%	
50) Toluene-d8	8.71	98	647182	45.48	ug/l	0.00
Spiked Amount			Recovery	=	90.96%	
62) 4-Bromofluorobenzene	11.13	95	214931	43.11	ug/l	0.00
Spiked Amount			Recovery	=	86.22%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX031919\
Data File : VX008185.D
Acq On : 19 Mar 2019 12:23
Operator : JC/SP
Sample : IBLK
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Mar 20 07:22:59 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X031919W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 19 11:34:11 2019
Response via : Initial Calibration

