

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031919\
 Data File : VX008197.D
 Acq On : 19 Mar 2019 18:05
 Operator : JC/SP
 Sample : K1860-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE120D1-20190311

Quant Time: Mar 20 08:01:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X031919W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 20 06:51:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	348275	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	558182	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	477935	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	194208	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	238118	45.40	ug/l	0.00
Spiked Amount			Recovery	=	90.80%	
35) Dibromofluoromethane	5.50	113	176209	47.16	ug/l	0.00
Spiked Amount			Recovery	=	94.32%	
50) Toluene-d8	8.71	98	692130	48.86	ug/l	0.00
Spiked Amount			Recovery	=	97.72%	
62) 4-Bromofluorobenzene	11.13	95	224180	45.96	ug/l	0.00
Spiked Amount			Recovery	=	91.92%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) 1,1,2-Trichlorotrifluoroet	2.39	101	46769	10.449	ug/l	91
12) 1,1-Dichloroethene	2.38	96	23628	5.591	ug/l	90
16) Acetone	2.45	43	3384	1.137	ug/l #	87
24) 1,1-Dichloroethane	3.70	63	9406	1.015	ug/l	98
27) cis-1,2-Dichloroethene	4.60	96	11519	2.213	ug/l	95
30) Chloroform	5.21	83	3417	0.394	ug/l	86
38) Carbon Tetrachloride	5.80	117	2203	0.380	ug/l #	82
44) Trichloroethene	7.22	130	3299041	701.459	ug/l	90
49) 1,4-Dioxane	7.74	88	2593	22.443	ug/l	94
55) 1,1,2-Trichloroethane	9.21	97	2801	0.608	ug/l	95
64) Tetrachloroethene	9.33	164	10421	2.417	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX031919\
Data File : VX008197.D
Acq On : 19 Mar 2019 18:05
Operator : JC/SP
Sample : K1860-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE120D1-20190311

Quant Time: Mar 20 08:01:06 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X031919W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 20 06:51:58 2019
Response via : Initial Calibration

