

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX031919\
 Data File : VX008198.D
 Acq On : 19 Mar 2019 18:30
 Operator : JC/SP
 Sample : K1860-04DL 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE120D1-20190311DL

Quant Time: Mar 26 02:53:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X031919W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 20 06:51:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	346046	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	556774	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	478304	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	196814	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	239750	46.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.00%	
35) Dibromofluoromethane	5.50	113	175422	47.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.12%	
50) Toluene-d8	8.71	98	694659	49.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.32%	
62) 4-Bromofluorobenzene	11.13	95	225333	46.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.62%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.39	101	5307	1.19	ug/l	89
12) 1,1-Dichloroethene	2.38	96	2716	0.65	ug/l	87
44) Trichloroethene	7.21	130	334517	71.31	ug/l	90
64) Tetrachloroethene	9.33	164	1361	0.32	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX031919\
Data File : VX008198.D
Acq On : 19 Mar 2019 18:30
Operator : JC/SP
Sample : K1860-04DL 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE120D1-20190311DL

Quant Time: Mar 26 02:53:12 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X031919W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 20 06:51:58 2019
Response via : Initial Calibration

