

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX031919\
 Data File : VX008203.D
 Acq On : 19 Mar 2019 20:34
 Operator : JC/SP
 Sample : K1882-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPS1-TT-MW305S-20190312

Quant Time: Mar 25 17:20:49 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X031919W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 20 06:51:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	335393	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	545610	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	459857	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	185301	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	234939	46.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.02%	
35) Dibromofluoromethane	5.50	113	170808	46.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.52%	
50) Toluene-d8	8.71	98	676238	48.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.66%	
62) 4-Bromofluorobenzene	11.13	95	215680	45.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.46%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX031919\
Data File : VX008203.D
Acq On : 19 Mar 2019 20:34
Operator : JC/SP
Sample : K1882-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW305S-20190312

Quant Time: Mar 25 17:20:49 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X031919W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 20 06:51:58 2019
Response via : Initial Calibration

