

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031919\
 Data File : VX008207.D
 Acq On : 19 Mar 2019 23:03
 Operator : JC/SP
 Sample : VX0319WBL02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0319WBL02

Quant Time: Mar 20 08:29:33 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X031919W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 20 06:51:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	342206	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	551073	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	475745	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	193786	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	236982	45.98	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.96%	
35) Dibromofluoromethane	5.50	113	171590	46.51	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.02%	
50) Toluene-d8	8.71	98	684623	48.95	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.90%	
62) 4-Bromofluorobenzene	11.13	95	222278	46.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.32%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX031919\
Data File : VX008207.D
Acq On : 19 Mar 2019 23:03
Operator : JC/SP
Sample : VX0319WBL02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0319WBL02

Quant Time: Mar 20 08:29:33 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X031919W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 20 06:51:58 2019
Response via : Initial Calibration

