

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031919\
 Data File : VX008211.D
 Acq On : 20 Mar 2019 00:43
 Operator : JC/SP
 Sample : K1882-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TB02-20190312

Quant Time: Mar 20 08:35:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X031919W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 20 06:51:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	335620	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	546011	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	460683	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	185129	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	233606	46.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.44%	
35) Dibromofluoromethane	5.50	113	172177	47.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.20%	
50) Toluene-d8	8.71	98	668242	48.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.44%	
62) 4-Bromofluorobenzene	11.13	95	213285	44.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.40%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX031919\
Data File : VX008211.D
Acq On : 20 Mar 2019 00:43
Operator : JC/SP
Sample : K1882-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TB02-20190312

Quant Time: Mar 20 08:35:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X031919W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 20 06:51:58 2019
Response via : Initial Calibration

