

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031919\
 Data File : VX008223.D
 Acq On : 20 Mar 2019 05:42
 Operator : JC/SP
 Sample : K1912-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 M3-T0-OZ(PRE)

Quant Time: Mar 20 09:12:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X031919W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 20 06:51:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	329771	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	530520	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	441958	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	168768	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	226111	45.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.06%	
35) Dibromofluoromethane	5.50	113	170191	47.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.84%	
50) Toluene-d8	8.71	98	654257	48.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.18%	
62) 4-Bromofluorobenzene	11.13	95	200780	43.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.62%	
Target Compounds						
19) Methyl tert-butyl Ether	3.20	73	35875	2.480	ug/l	98
24) 1,1-Dichloroethane	3.71	63	9740	1.110	ug/l #	94
25) 2-Butanone	4.69	43	10349	2.489	ug/l	91
27) cis-1,2-Dichloroethene	4.61	96	14292	2.900	ug/l	98
29) Tetrahydrofuran	5.14	42	39507	15.199	ug/l	99
44) Trichloroethene	7.21	130	698542	156.272	ug/l	88
64) Tetrachloroethene	9.34	164	30731	7.707	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX031919\
Data File : VX008223.D
Acq On : 20 Mar 2019 05:42
Operator : JC/SP
Sample : K1912-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
M3-T0-OZ(PRE)

Quant Time: Mar 20 09:12:40 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X031919W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 20 06:51:58 2019
Response via : Initial Calibration

