

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031919\
 Data File : VX008188.D
 Acq On : 19 Mar 2019 13:44
 Operator : JC/SP
 Sample : VX0319WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0319WBL01

Quant Time: Mar 20 07:33:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X031919W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 20 06:51:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	361204	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	579972	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	485633	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	194223	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	246398	45.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.58%	
35) Dibromofluoromethane	5.50	113	182452	46.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.98%	
50) Toluene-d8	8.71	98	719671	48.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.78%	
62) 4-Bromofluorobenzene	11.13	95	228290	45.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.08%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX031919\
Data File : VX008188.D
Acq On : 19 Mar 2019 13:44
Operator : JC/SP
Sample : VX0319WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0319WBL01

Quant Time: Mar 20 07:33:07 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X031919W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 20 06:51:58 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.67 min Scan# 740

Delta R.T. -0.00 min

Lab File: VX008188.D

Acq: 19 Mar 2019 13:44

Instrument :

MSVOA_X

ClientSampleId :


VX0319WBL01

Tot Ion:168 Resp: 361204

Ion Ratio Lower Upper

168 100

99 57.1 37.6 56.4#



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.67



150000

100000

50000

0

Time--> 5.50 5.60 5.70 5.80



#33

1,2-Dichloroethane-d4

Concen: 45.293 ug/l

RT: 6.06 min Scan# 805

Delta R.T. -0.01 min

Lab File: VX008188.D


Acq: 19 Mar 2019 13:44

Tgt Ion: 65 Resp: 246398

Ion Ratio Lower Upper

65 100

67 54.0 0.0 110.0



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.06

100000

80000


60000


40000

20000

0

Time--> 5.90 6.00 6.10 6.20





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.86 min Scan# 936

Delta R.T. -0.00 min

Lab File: VX008188.D

Acq: 19 Mar 2019 13:44

Instrument :

MSVOA_X

ClientSampleId :

VX0319WBL01


Tot Ion:114 Resp: 579972

Ion Ratio Lower Upper

114 100

63 24.0 0.0 39.8




88 16.8 0.0 31.4



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 46.993 ug/l

RT: 5.50 min Scan# 713

Delta R.T. -0.00 min

Lab File: VX008188.D

Acq: 19 Mar 2019 13:44


Tgt Ion:113 Resp: 182452

Ion Ratio Lower Upper

113 100

111 103.0 81.4 122.0



192 20.4 19.4 29.0




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 48.891 ug/l

RT: 8.71 min Scan# 1240

Delta R.T. -0.00 min

Lab File: VX008188.D

Acq: 19 Mar 2019 13:44

Instrument :

MSVOA_X

ClientSampled :


VX0319WBL01

Tot Ion: 98 Resp: 719671

Ion Ratio Lower Upper




98 100

100 62.8 51.7 77.5



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): V



#62

4-Bromofluorobenzene

Concen: 45.042 ug/l

RT: 11.13 min Scan# 1637

Delta R.T. -0.00 min

Lab File: VX008188.D

Acq: 19 Mar 2019 13:44


Tgt Ion: 95 Resp: 228290

Ion Ratio Lower Upper

95 100

174 77.1 0.0 182.8



176 74.7 0.0 178.0




Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1469
Delta R.T. -0.00 min
Lab File: VX008188.D
Acq: 19 Mar 2019 13:44

Instrument :
MSVOA_X
ClientSampleId :
VX0319WBL01

Tot Ion	Ratio	Lower	Upper
117	100		
82	61.9	44.3	66.5
119	32.4	25.3	37.9



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.08 min Scan# 1792
Delta R.T. -0.00 min
Lab File: VX008188.D
Acq: 19 Mar 2019 13:44

Tgt Ion	Ratio	Lower	Upper
152	100		
115	60.1	38.6	115.7
150	157.7	0.0	348.4

