

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031922\
 Data File : VX027555.D
 Acq On : 19 Mar 2022 13:12
 Operator : JC/MD
 Sample : N2000-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-20220315

Quant Time: Mar 21 01:21:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 15:02:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	354037	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	582046	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	529170	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	243373	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	192201	44.069	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	88.140%
35) Dibromofluoromethane	5.391	113	187008	49.938	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.880%
50) Toluene-d8	8.653	98	678110	47.853	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.700%
62) 4-Bromofluorobenzene	11.085	95	235235	45.164	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	90.320%
Target Compounds						
16) Acetone	2.386	43	4879	2.745	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031922\
Data File : VX027555.D
Acq On : 19 Mar 2022 13:12
Operator : JC/MD
Sample : N2000-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-20220315

Quant Time: Mar 21 01:21:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 15:02:59 2022
Response via : Initial Calibration

