

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031922\
 Data File : VX027558.D
 Acq On : 19 Mar 2022 14:23
 Operator : JC/MD
 Sample : N2000-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE-132D6-20220315

Quant Time: Mar 21 01:21:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 15:02:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	351550	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	579665	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	536761	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	252926	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	190225	43.924	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	87.840%		
35) Dibromofluoromethane	5.391	113	185061	49.621	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	99.240%		
50) Toluene-d8	8.653	98	675167	47.841	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	95.680%		
62) 4-Bromofluorobenzene	11.085	95	245915	47.408	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	94.820%		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	1103	0.327	ug/l	98
9) 1,1,2-Trichlorotrifluo...	2.331	101	86690	25.480	ug/l	92
12) 1,1-Dichloroethene	2.318	96	27179	8.326	ug/l	94
16) Acetone	2.379	43	3133	1.775	ug/l	94
24) 1,1-Dichloroethane	3.611	63	9189	1.515	ug/l	99
27) cis-1,2-Dichloroethene	4.495	96	15099	3.739	ug/l	85
30) Chloroform	5.092	83	6460	0.965	ug/l	97
38) Carbon Tetrachloride	5.690	117	6001	1.029	ug/l	97
42) 1,2-Dichloroethane	6.098	62	1495	0.273	ug/l	100
44) Trichloroethene	7.135	130	5968084	1344.583	ug/l	100
49) 1,4-Dioxane	7.671	88	1734	16.691	ug/l	92
55) 1,1,2-Trichloroethane	9.152	97	3860	0.968	ug/l	91
64) Tetrachloroethene	9.274	164	17982	4.309	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031922\
Data File : VX027558.D
Acq On : 19 Mar 2022 14:23
Operator : JC/MD
Sample : N2000-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE-132D6-20220315

Quant Time: Mar 21 01:21:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 15:02:59 2022
Response via : Initial Calibration

