

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031922\
 Data File : VX027566.D
 Acq On : 19 Mar 2022 17:32
 Operator : JC/MD
 Sample : N2000-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE-139D1-20220316

Quant Time: Mar 21 01:24:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 15:02:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	330371	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	545000	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	475855	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	214915	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	178499	43.859	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	87.720%
35) Dibromofluoromethane	5.391	113	176280	50.273	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.540%
50) Toluene-d8	8.653	98	630135	47.490	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	94.980%
62) 4-Bromofluorobenzene	11.085	95	213577	43.793	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	87.580%

Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.331	101	16652	5.208	ug/l	96
12) 1,1-Dichloroethene	2.319	96	10172	3.316	ug/l	98
16) Acetone	2.380	43	3824	2.305	ug/l #	85
24) 1,1-Dichloroethane	3.617	63	3146	0.552	ug/l #	92
27) cis-1,2-Dichloroethene	4.495	96	6007	1.583	ug/l	82
30) Chloroform	5.099	83	4235	0.673	ug/l	94
38) Carbon Tetrachloride	5.678	117	6059	1.105	ug/l	87
44) Trichloroethene	7.129	130	679550	162.838	ug/l	100
49) 1,4-Dioxane	7.684	88	540	5.528	ug/l #	81
55) 1,1,2-Trichloroethane	9.159	97	3597	0.959	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031922\
Data File : VX027566.D
Acq On : 19 Mar 2022 17:32
Operator : JC/MD
Sample : N2000-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE-139D1-20220316

Quant Time: Mar 21 01:24:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 15:02:59 2022
Response via : Initial Calibration

