

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031922\
 Data File : VX027569.D
 Acq On : 19 Mar 2022 18:43
 Operator : JC/MD
 Sample : N2000-19
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE-105D1-20220316

Quant Time: Mar 21 01:25:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 15:02:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	329754	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	548192	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	481574	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	220614	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	194005	47.76	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.52%
35) Dibromofluoromethane	5.391	113	178987	50.75	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.50%
50) Toluene-d8	8.653	98	634047	47.51	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.02%
62) 4-Bromofluorobenzene	11.085	95	220520	44.95	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	89.90%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	1159	0.37	ug/l	88
9) 1,1,2-Trichlorotrifluo...	2.325	101	13333	4.18	ug/l	89
12) 1,1-Dichloroethene	2.319	96	2192	0.72	ug/l	91
27) cis-1,2-Dichloroethene	4.495	96	3235	0.85	ug/l	93
44) Trichloroethene	7.129	130	346535	82.56	ug/l	99
49) 1,4-Dioxane	7.671	88	1273	12.96	ug/l	91
64) Tetrachloroethene	9.275	164	1458	0.39	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031922\
Data File : VX027569.D
Acq On : 19 Mar 2022 18:43
Operator : JC/MD
Sample : N2000-19
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE-105D1-20220316

Quant Time: Mar 21 01:25:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 15:02:59 2022
Response via : Initial Calibration

