

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031924\
 Data File : VX040713.D
 Acq On : 19 Mar 2024 11:22
 Operator : JC/MD
 Sample : P1788-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PEZOMETER-200

Quant Time: Mar 20 04:33:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X030524W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Mar 05 13:21:29 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.891	128	15957	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.756	114	92348	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	86672	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.952	65	49737	31.209	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	104.033%
60) 4-Bromofluorobenzene	11.079	95	43568	29.305	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	97.700%
63) Toluene-d8	8.646	98	109618	28.619	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	95.400%
Target Compounds						
					Qvalue	
15) Acetone	2.385	58	695	3.939	ug/l #	65

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031924\
Data File : VX040713.D
Acq On : 19 Mar 2024 11:22
Operator : JC/MD
Sample : P1788-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PEZOMETER-200

Quant Time: Mar 20 04:33:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X030524W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Mar 05 13:21:29 2024
Response via : Initial Calibration

