

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031924\
 Data File : VX040714.D
 Acq On : 19 Mar 2024 11:44
 Operator : JC/MD
 Sample : P1788-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PEZOMETER-206

Quant Time: Mar 20 04:33:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X030524W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Mar 05 13:21:29 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.903	128	15208	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	82463	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	77987	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.952	65	44588	29.356	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	97.867%
60) 4-Bromofluorobenzene	11.079	95	38170	28.534	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	95.100%
63) Toluene-d8	8.647	98	99585	28.895	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	96.300%
Target Compounds						
					Qvalue	
15) Acetone	2.373	58	685	4.074	ug/l	# 64
16) Carbon Disulfide	2.520	76	550	0.221	ug/l	95
25) Chloroform	5.099	83	5825	2.763	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031924\
Data File : VX040714.D
Acq On : 19 Mar 2024 11:44
Operator : JC/MD
Sample : P1788-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PEZOMETER-206

Quant Time: Mar 20 04:33:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X030524W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Mar 05 13:21:29 2024
Response via : Initial Calibration

