

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031925\
 Data File : VX045340.D
 Acq On : 19 Mar 2025 10:41
 Operator : JC/MD
 Sample : VSTDCCC020
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC020

Quant Time: Mar 20 01:30:50 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X022125W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Feb 22 01:06:36 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	89	0.01
2 M	Dichlorodifluoromethane	2.529	2.632	-4.1	95	0.00
3 M	Chloromethane	3.029	2.718	10.3	81	0.01
4 M	Vinyl Chloride	2.923	2.544	13.0	80	0.00
5 M	Bromomethane	1.017	1.128	-10.9	103	0.00
6 M	Chloroethane	1.583	1.341	15.3	72	0.00
7 M	Trichlorofluoromethane	3.928	3.690	6.1	89	0.00
8 T	Diethyl Ether	1.486	1.267	14.7	79	0.00
9	1,1,2-Trichlorotrifluoroeth	2.333	2.471	-5.9	100	0.00
10 M	1,1-Dichloroethene	2.279	2.299	-0.9	96	0.00
11	Methyl Iodide	2.579	2.702	-4.8	111	0.00
12	Methyl Acetate	2.646	4.412	-66.7#	159#	0.00
13 M	Acrolein	0.530	0.631	-19.1	119	0.00
14 M	Acrylonitrile	1.293	1.467	-13.5	105	0.00
15 M	Acetone	0.355	0.395	-11.3	100	0.00
16 M	Carbon Disulfide	6.464	5.226	19.2	78	0.00
17	Allyl chloride	4.262	4.176	2.0	94	0.00
18 M	Methylene Chloride	2.563	2.580	-0.7	93	0.00
19 M	trans-1,2-Dichloroethene	2.306	2.205	4.4	90	0.00
20 T	Diisopropyl ether	8.191	8.189	0.0	94	0.00
21 M	1,1-Dichloroethane	4.555	4.617	-1.4	95	0.00
22 M	cis-1,2-Dichloroethene	2.763	2.788	-0.9	94	0.00
23 M	tert-Butyl Alcohol	0.451	0.427	5.3	90	0.00
24 M	Methyl tert-Butyl Ether	7.438	7.562	-1.7	96	0.00
25 M	Chloroform	4.400	4.726	-7.4	99	0.00
26	Cyclohexane	4.205	4.013	4.6	89	0.00
27 s	1,2-Dichloroethane-d4	2.914	3.187	-9.4	99	0.00
28 I	1,4-Difluorobenzene	1.000	1.000	0.0	85	0.00
29	1,1-Dichloropropene	0.552	0.569	-3.1	90	0.00
30 M	2-Butanone	0.319	0.396	-24.1	109	0.00
31	2,2-Dichloropropane	0.624	0.597	4.3	87	0.00
32 M	1,1,1-Trichloroethane	0.669	0.727	-8.7	96	0.00
33 M	Carbon Tetrachloride	0.564	0.602	-6.7	94	0.00
34 M	Benzene	1.731	1.833	-5.9	92	0.00
35	Methacrylonitrile	0.341	0.411	-20.5	106	0.00
36 M	1,2-Dichloroethane	0.592	0.694	-17.2	102	0.00
37 M	Trichloroethene	0.405	0.424	-4.7	91	0.00
38	Methylcyclohexane	0.741	0.739	0.3	90	0.00
39 M	1,2-Dichloropropane	0.430	0.448	-4.2	90	0.00
40	Dibromomethane	0.309	0.346	-12.0	98	0.00
41 M	Bromodichloromethane	0.614	0.683	-11.2	98	0.00
42 M	Vinyl Acetate	1.237	1.256	-1.5	90	0.00
43	Ethyl Acetate	0.598	0.716	-19.7	106	0.00
44	Isopropyl Acetate	0.984	1.085	-10.3	101	0.00
45 T	1,4-Dioxane	0.010	0.010	0.0	83	0.00
46	Methyl methacrylate	0.482	0.577	-19.7	108	0.00
47	n-amyl Acetate	0.854	0.885	-3.6	94	0.00
48 M	t-1,3-Dichloropropene	0.591	0.563	4.7	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031925\
 Data File : VX045340.D
 Acq On : 19 Mar 2025 10:41
 Operator : JC/MD
 Sample : VSTDCCC020
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC020

Quant Time: Mar 20 01:30:50 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X022125W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Feb 22 01:06:36 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	cis-1,3-Dichloropropene	0.670	0.674	-0.6	91	0.00
50 M	1,1,2-Trichloroethane	0.403	0.444	-10.2	94	0.00
51	Ethyl methacrylate	0.633	0.676	-6.8	99	0.00
52	1,3-Dichloropropane	0.706	0.796	-12.7	97	0.00
53 M	Dibromochloromethane	0.445	0.483	-8.5	95	0.00
54 M	1,2-Dibromoethane	0.409	0.450	-10.0	96	0.00
55 M	2-Chloroethyl vinyl ether	0.325	0.336	-3.4	91	0.00
56 M	Bromoform	0.279	0.296	-6.1	98	0.00
57 I	Chlorobenzene-d5	1.000	1.000	0.0	88	0.00
58 M	4-Methyl-2-Pentanone	0.682	0.832	-22.0	108	0.00
59 M	2-Hexanone	0.493	0.608	-23.3	110	0.00
60 S	4-Bromofluorobenzene	0.561	0.584	-4.1	92	0.00
61 M	Tetrachloroethene	0.369	0.402	-8.9	96	0.00
62 M	Toluene	1.987	2.063	-3.8	93	0.00
63 S	Toluene-d8	1.660	1.665	-0.3	88	0.00
64 M	Chlorobenzene	1.231	1.280	-4.0	93	0.00
65	1,1,1,2-Tetrachloroethane	0.406	0.428	-5.4	96	0.00
66 M	Ethyl Benzene	2.176	2.264	-4.0	95	0.00
67 M	m/p-Xylenes	0.811	0.854	-5.3	92	0.00
68 M	o-Xylene	0.795	0.811	-2.0	90	0.00
69 M	Styrene	1.328	1.394	-5.0	93	0.00
70	Isopropylbenzene	2.053	2.224	-8.3	96	0.00
71 M	1,1,2,2-Tetrachloroethane	0.675	0.777	-15.1	103	0.00
72	1,2,3-Trichloropropane	0.555	0.644	-16.0	105	0.00
73	Bromobenzene	0.473	0.488	-3.2	92	0.00
74	n-propylbenzene	2.443	2.564	-5.0	95	0.00
75	2-Chlorotoluene	1.477	1.596	-8.1	96	0.00
76	1,3,5-Trimethylbenzene	1.693	1.806	-6.7	94	0.00
77	t-1,4-Dichloro-2-butene	0.187	0.180	3.7	99	0.00
78	4-Chlorotoluene	1.646	1.755	-6.6	96	0.00
79	tert-butylbenzene	1.706	1.798	-5.4	93	0.00
80	1,2,4-Trimethylbenzene	1.696	1.802	-6.3	93	0.00
81	sec-Butylbenzene	2.141	2.320	-8.4	97	0.00
82	p-Isopropyltoluene	1.735	1.845	-6.3	95	0.00
83 M	1,3-Dichlorobenzene	0.879	0.891	-1.4	90	0.00
84 M	1,4-Dichlorobenzene	0.867	0.904	-4.3	93	0.00
85	n-Butylbenzene	1.584	1.603	-1.2	93	0.00
86 T	Hexachloroethane	0.306	0.307	-0.3	96	0.00
87 M	1,2-Dichlorobenzene	0.863	0.898	-4.1	91	0.00
88	1,2-Dibromo-3-Chloropropane	0.131	0.150	-14.5	109	0.00
89	1,2,4-Trichlorobenzene	0.523	0.502	4.0	90	0.00
90	Hexachlorobutadiene	0.213	0.218	-2.3	92	0.00
91 M	Naphthalene	1.843	1.848	-0.3	92	0.00
92	1,2,3-Trichlorobenzene	0.538	0.529	1.7	89	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0