

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX032019\
 Data File : VX008236.D
 Acq On : 20 Mar 2019 12:45
 Operator : JC/SP
 Sample : K1860-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-02-20190311

Quant Time: Mar 22 06:44:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X031919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 13:15:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	343818	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	562147	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	474202	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	192297	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	238934	46.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.28%	
35) Dibromofluoromethane	5.50	113	177575	47.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.38%	
50) Toluene-d8	8.71	98	690400	48.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.78%	
62) 4-Bromofluorobenzene	11.13	95	219468	44.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.34%	
Target Compounds						
16) Acetone	2.45	43	14231	4.846	ug/l	96
52) Toluene	8.79	92	22120	1.925	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX032019\
Data File : VX008236.D
Acq On : 20 Mar 2019 12:45
Operator : JC/SP
Sample : K1860-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
EB-02-20190311

Quant Time: Mar 22 06:44:31 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X031919W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 19 13:15:13 2019
Response via : Initial Calibration

