

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX032019\
 Data File : VX008247.D
 Acq On : 20 Mar 2019 17:19
 Operator : JC/SP
 Sample : K1963-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPS1-TT-MW307I-20190314

Quant Time: Mar 22 07:16:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X031919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 13:15:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	324935	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	533641	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	454224	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	185678	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	228599	46.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.42%	
35) Dibromofluoromethane	5.50	113	169621	47.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.96%	
50) Toluene-d8	8.71	98	657578	48.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.10%	
62) 4-Bromofluorobenzene	11.13	95	211878	45.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.86%	
Target Compounds						
44) Trichloroethene	7.21	130	3904	0.868	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX032019\
Data File : VX008247.D
Acq On : 20 Mar 2019 17:19
Operator : JC/SP
Sample : K1963-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW307I-20190314

Quant Time: Mar 22 07:16:09 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X031919W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 19 13:15:13 2019
Response via : Initial Calibration

