

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX032019\
 Data File : VX008262.D
 Acq On : 21 Mar 2019 00:22
 Operator : JC/SP
 Sample : K1900-13DL 20X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPS1-TT-MW305I-20190313DL

Quant Time: Mar 22 07:47:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X031919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 13:15:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	344049	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	558334	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	477023	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	193507	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	237231	45.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.56%	
35) Dibromofluoromethane	5.50	113	176310	47.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.34%	
50) Toluene-d8	8.71	98	694072	48.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.96%	
62) 4-Bromofluorobenzene	11.13	95	223576	45.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.64%	
Target Compounds						
44) Trichloroethene	7.21	130	286309	60.860	ug/l	Qvalue 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX032019\
Data File : VX008262.D
Acq On : 21 Mar 2019 00:22
Operator : JC/SP
Sample : K1900-13DL 20X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW305I-20190313DL

Quant Time: Mar 22 07:47:39 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X031919W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 19 13:15:13 2019
Response via : Initial Calibration

