

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX032019\
 Data File : VX008268.D
 Acq On : 21 Mar 2019 02:51
 Operator : JC/SP
 Sample : K1912-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 M3-T2-OZ(PRE)

Quant Time: Mar 22 07:59:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X031919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 13:15:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	322726	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	527460	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	450061	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	179260	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	228581	47.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.06%	
35) Dibromofluoromethane	5.50	113	168293	47.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.32%	
50) Toluene-d8	8.71	98	652401	48.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.46%	
62) 4-Bromofluorobenzene	11.13	95	207015	44.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.82%	
Target Compounds						
24) 1,1-Dichloroethane	3.70	63	6433	0.749	ug/l	98
32) 1,1,1-Trichloroethane	5.50	97	2732	0.409	ug/l #	47
44) Trichloroethene	7.21	130	72520	16.318	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX032019\
Data File : VX008268.D
Acq On : 21 Mar 2019 02:51
Operator : JC/SP
Sample : K1912-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
M3-T2-OZ(PRE)

Quant Time: Mar 22 07:59:52 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X031919W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 19 13:15:13 2019
Response via : Initial Calibration

