

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX032019\
 Data File : VX008270.D
 Acq On : 21 Mar 2019 03:41
 Operator : JC/SP
 Sample : K1912-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 M3-T4-OZ(POST-HYPO)

Quant Time: Mar 22 08:04:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X031919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 13:15:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	61079	50.00	ug/l	-0.04
34) 1,4-Difluorobenzene	6.84	114	29062	50.00	ug/l	-0.02
63) Chlorobenzene-d5	10.11	117	116143	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	92928	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.02	65	2170	2.36	ug/l	-0.04
Spiked Amount						
			Recovery	=		4.72%
35) Dibromofluoromethane	5.46	113	1000	5.14	ug/l	-0.04
Spiked Amount						
			Recovery	=		10.28%
50) Toluene-d8	8.71	98	136961	185.68	ug/l	0.00
Spiked Amount						
			Recovery	=		371.36%
62) 4-Bromofluorobenzene	11.13	95	85425	336.35	ug/l	0.00
Spiked Amount						
			Recovery	=		672.70%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX032019\
Data File : VX008270.D
Acq On : 21 Mar 2019 03:41
Operator : JC/SP
Sample : K1912-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
M3-T4-OZ(POST-HYPO)

Quant Time: Mar 22 08:04:18 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X031919W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 19 13:15:13 2019
Response via : Initial Calibration

