

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX032019\
 Data File : VX008274.D
 Acq On : 21 Mar 2019 05:21
 Operator : JC/SP
 Sample : K1912-14
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 M3-T2-UV(PRE)

Quant Time: Mar 22 08:15:32 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X031919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 13:15:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	290971	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	488789	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	407594	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	163896	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	207859	47.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.86%	
35) Dibromofluoromethane	5.50	113	160388	49.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.04%	
50) Toluene-d8	8.71	98	590853	47.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.26%	
62) 4-Bromofluorobenzene	11.13	95	187198	43.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.64%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluoroet	2.39	101	3868	1.034	ug/l	92
16) Acetone	2.45	43	5281	2.125	ug/l #	84
24) 1,1-Dichloroethane	3.70	63	7287	0.941	ug/l	96
32) 1,1,1-Trichloroethane	5.49	97	2818	0.468	ug/l #	33
44) Trichloroethene	7.21	130	5411	1.314	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX032019\
Data File : VX008274.D
Acq On : 21 Mar 2019 05:21
Operator : JC/SP
Sample : K1912-14
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
M3-T2-UV(PRE)

Quant Time: Mar 22 08:15:32 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X031919W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 19 13:15:13 2019
Response via : Initial Calibration

