

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX032019\
 Data File : VX008248.D
 Acq On : 20 Mar 2019 17:44
 Operator : JC/SP
 Sample : K1963-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-DUP03-20190314

Quant Time: Mar 22 07:17:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X031919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 13:15:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	321476	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	524703	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	441839	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	178292	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	228591	47.21	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.42%	
35) Dibromofluoromethane	5.50	113	167243	47.61	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.22%	
50) Toluene-d8	8.71	98	645564	48.48	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.96%	
62) 4-Bromofluorobenzene	11.13	95	203763	44.44	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.88%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX032019\
Data File : VX008248.D
Acq On : 20 Mar 2019 17:44
Operator : JC/SP
Sample : K1963-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-DUP03-20190314

Quant Time: Mar 22 07:17:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X031919W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 19 13:15:13 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.67 min Scan# 740

Delta R.T. 0.00 min

Lab File: VX008248.D

Acq: 20 Mar 2019 17:44

Instrument :

MSVOA_X

ClientSampleId :


BP-DUP03-20190314

Tot Ion:168 Resp: 321476

Ion Ratio Lower Upper

168 100

99 56.2 37.6 56.4



Abundance Ion 167.90 (167.60 to 168.60): VX008248.D (-726) (-)

Ion 98.90 (98.60 to 99.60): VX008248.D (-726) (-)



5.67

100000

50000

0

Time--> 5.50 5.60 5.70 5.80



#33

1,2-Dichloroethane-d4

Concen: 47.213 ug/l

RT: 6.07 min Scan# 806

Delta R.T. 0.00 min

Lab File: VX008248.D


Acq: 20 Mar 2019 17:44

Tgt Ion: 65 Resp: 228591

Ion Ratio Lower Upper

65 100

67 52.4 0.0 110.0



Abundance Ion 64.90 (64.60 to 65.60): VX008248.D (-794) (-)

Ion 66.90 (66.60 to 67.60): VX008248.D (-794) (-)

6.07

100000

80000


60000


40000

20000

0

Time--> 5.90 6.00 6.10 6.20





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.86 min Scan# 936

Delta R.T. 0.00 min

Lab File: VX008248.D

Acq: 20 Mar 2019 17:44

Instrument :

MSVOA_X

ClientSampleId :

BP-DUP03-20190314


Tot Ion:114 Resp: 524703

Ion Ratio Lower Upper

114 100

63 23.0 0.0 39.8




88 17.0 0.0 31.4



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 47.613 ug/l

RT: 5.50 min Scan# 713

Delta R.T. -0.00 min

Lab File: VX008248.D

Acq: 20 Mar 2019 17:44


Tgt Ion:113 Resp: 167243

Ion Ratio Lower Upper

113 100

111 101.3 81.4 122.0



192 20.4 19.4 29.0




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 48.476 ug/l

RT: 8.71 min Scan# 1240

Delta R.T. 0.00 min

Lab File: VX008248.D

Acq: 20 Mar 2019 17:44

Instrument :

MSVOA_X

ClientSampleId :





BP-DUP03-20190314

Tot Ion: 98 Resp: 645564

Ion Ratio Lower Upper

98 100

100 63.2 51.7 77.5



#62

4-Bromofluorobenzene

Concen: 44.437 ug/l

RT: 11.13 min Scan# 1637

Delta R.T. 0.00 min

Lab File: VX008248.D

Acq: 20 Mar 2019 17:44




Tgt Ion: 95 Resp: 203763


Ion Ratio Lower Upper

95 100

174 78.7 0.0 182.8

176 75.7 0.0 178.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1469
Delta R.T. 0.00 min
Lab File: VX008248.D
Acq: 20 Mar 2019 17:44

Instrument :
MSVOA_X
ClientSampleId :
BP-DUP03-20190314

Tot Ion:117 Resp: 441839
Ion Ratio Lower Upper
117 100
82 60.2 44.3 66.5
119 31.6 25.3 37.9



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.08 min Scan# 1792
Delta R.T. 0.00 min
Lab File: VX008248.D
Acq: 20 Mar 2019 17:44

Tgt Ion:152 Resp: 178292
Ion Ratio Lower Upper
152 100
115 59.7 38.6 115.7
150 156.3 0.0 348.4

