

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032023\
 Data File : VX034626.D
 Acq On : 20 Mar 2023 14:35
 Operator : JC/MD
 Sample : 01828-01
 Misc : 5.06g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 AVENEL-60

Quant Time: Mar 21 01:42:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	231579	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	431239	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	340442	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	161978	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	159202	47.255	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	94.500%		
35) Dibromofluoromethane	5.385	113	117057	41.112	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	82.220%		
50) Toluene-d8	8.647	98	487066	45.824	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	91.640%#		
62) 4-Bromofluorobenzene	11.085	95	167592	39.036	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	78.080%#		
Target Compounds						
16) Acetone	2.416	43	5653	3.562	ug/l	98
18) Methyl Acetate	2.715	43	136340	26.293	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032023\
Data File : VX034626.D
Acq On : 20 Mar 2023 14:35
Operator : JC/MD
Sample : 01828-01
Misc : 5.06g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
AVENEL-60

Quant Time: Mar 21 01:42:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

