

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032023\
 Data File : VX034636.D
 Acq On : 20 Mar 2023 18:30
 Operator : JC/MD
 Sample : 02012-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-2023-3-19

Quant Time: Mar 21 01:46:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	279952	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	469669	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	416145	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	178807	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	204372	50.181	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.360%
35) Dibromofluoromethane	5.385	113	147523	47.573	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.140%
50) Toluene-d8	8.647	98	562395	48.582	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.160%
62) 4-Bromofluorobenzene	11.079	95	222140	47.508	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032023\
Data File : VX034636.D
Acq On : 20 Mar 2023 18:30
Operator : JC/MD
Sample : 02012-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-2023-3-19

Quant Time: Mar 21 01:46:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

