

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032023\
 Data File : VX034640.D
 Acq On : 20 Mar 2023 20:04
 Operator : JC/MD
 Sample : 02012-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 B-23-GW01

Quant Time: Mar 21 01:47:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	280136	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	472032	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	418992	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	178277	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	207686	50.961	ug/l	0.01
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.920%
35) Dibromofluoromethane	5.385	113	147518	47.333	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.660%
50) Toluene-d8	8.647	98	564529	48.522	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.040%
62) 4-Bromofluorobenzene	11.079	95	224529	47.778	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.560%
Target Compounds						
						Qvalue
16) Acetone	2.379	43	9415	4.904	ug/l	# 84
18) Methyl Acetate	2.696	43	3157	0.503	ug/l	97
52) Toluene	8.720	92	113504	13.092	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032023\
Data File : VX034640.D
Acq On : 20 Mar 2023 20:04
Operator : JC/MD
Sample : 02012-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
B-23-GW01

Quant Time: Mar 21 01:47:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

