

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032023\
 Data File : VX034634.D
 Acq On : 20 Mar 2023 17:42
 Operator : JC/MD
 Sample : 02015-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Mar 21 01:45:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	255153	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.763	114	426786	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	419662	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	178688	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	169713	45.721	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.440%
35) Dibromofluoromethane	5.385	113	131213	46.565	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.140%
50) Toluene-d8	8.647	98	511879	48.661	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.320%
62) 4-Bromofluorobenzene	11.079	95	232496	54.718	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	109.440%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032023\
Data File : VX034634.D
Acq On : 20 Mar 2023 17:42
Operator : JC/MD
Sample : 02015-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Mar 21 01:45:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. 0.012 min
Lab File: VX034634.D
Acq: 20 Mar 2023 17:42




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4


Tgt Ion:168 Resp: 255153
Ion Ratio Lower Upper
168 100
99 58.8 45.7 68.5



#33
1,2-Dichloroethane-d4
Concen: 45.721 ug/l
RT: 5.952 min Scan# 798
Delta R.T. 0.006 min
Lab File: VX034634.D
Acq: 20 Mar 2023 17:42

Tgt Ion: 65 Resp: 169713
Ion Ratio Lower Upper
65 100
67 51.8 0.0 104.0









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.763 min Scan# 91
Delta R.T. 0.006 min
Lab File: VX034634.D
Acq: 20 Mar 2023 17:42




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4


Tgt Ion:114 Resp: 426786
Ion Ratio Lower Upper
114 100
63 22.5 0.0 42.2
88 16.6 0.0 33.2



#35
Dibromofluoromethane
Concen: 46.565 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.006 min
Lab File: VX034634.D
Acq: 20 Mar 2023 17:42

Tgt Ion:113 Resp: 131213
Ion Ratio Lower Upper
113 100
111 103.7 83.7 125.5
192 19.4 15.4 23.2









#50
Toluene-d8
Concen: 48.661 ug/l
RT: 8.647 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX034634.D
Acq: 20 Mar 2023 17:42




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4


Tgt Ion: 98 Resp: 511879
Ion Ratio Lower Upper
98 100
100 65.6 52.5 78.7



#62
4-Bromofluorobenzene
Concen: 54.718 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX034634.D
Acq: 20 Mar 2023 17:42

Tgt Ion: 95 Resp: 232496
Ion Ratio Lower Upper
95 100
174 69.2 0.0 147.4
176 67.8 0.0 140.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX034634.D
Acq: 20 Mar 2023 17:42

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Tgt Ion:117 Resp: 419662
Ion Ratio Lower Upper
117 100
82 59.6 45.0 67.4
119 32.8 26.0 39.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX034634.D
Acq: 20 Mar 2023 17:42

Tgt Ion:152 Resp: 178688
Ion Ratio Lower Upper
152 100
115 62.8 45.0 135.0
150 156.8 0.0 352.4

