



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032023\
Data File : VX034619.D
Acq On : 20 Mar 2023 09:22
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Title : SW846 8260
Last Update : Mon Mar 13 10:29:32 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.2	20184	PASS
75	95	30	60	52.8	52683	PASS
95	95	100	100	100.0	99851	PASS
96	95	5	9	6.6	6618	PASS
173	174	0.00	2	1.0	730	PASS
174	95	50	100	70.2	70136	PASS
175	174	5	9	7.9	5513	PASS
176	174	95	101	97.3	68232	PASS
177	176	5	9	6.6	4511	PASS